REF 6653

UNITS - 1

RX Only

Uni-lim™ *Coaxial Circuit*

60" Adult Coaxial Circle Anesthesia Breathing Circuit

LATEX Not Made With Natural Rubber Latex

DEHP Not Made With DEHP

CE 0482

CAUTION: Before use, follow the instructions provided by the manufacturer of the breathing machine for proper testing and installation of he circuit. Ensure all connections are secure. Use of sterilization or cleaning solutions may leave this device nonfunctional.

The Westmed coaxial re-breathing circuit is designed to function as a standard circle anesthesia breathing system. It is attached to the anesthesia machine as illustrated.
Note the inspiratory and expiratory connections.

IMPORTANT: Breathing system accessories, such as breathing filters and heat and moisture exchangers (HME) should be placed at the patient end of the circuit for maximum efficiency.



Westmed
5580 S. Nogales Highway
Tucson, Arizona 85706 USA
Phone: 800-975-7987
Fax: 520-294-6061
www.westmedinc.com

EC REP

MT Promedt Consulting GmbH
Altenhofstr. 80
66386 St. Ingbert, Germany

Patent Pending
Assembled in Mexico
LABEL PN 78837, Rev. 01

LOT