



A.C.E. Stat Check II CO₂ Indicator

Airway Circulation Evaluator



REF SC200

Cautions & Warnings:
THIS DEVICE IS NOT A SUBSTITUTE FOR OBSERVATION OF THE PATIENT. THE STAT-CHECK II MUST NOT BE RELIED UPON AS THE SOLE INDICATOR OF RESUSCITATION EFFECTIVENESS. THIS DEVICE IS AN INDICATOR OF BOTH VENTILATION AND CIRCULATION. THE INDICATOR DYE SHOULD BE CHECKED FREQUENTLY DURING USE. BRAIN DEATH OCCURS IN 4-6 MINUTES WITHOUT OXYGENATED BLOOD FLOW. THE SIGNS OF VENTILATION (i.e. BILATERAL BREATH-SOUND, "FOG SIGN", etc.) and CIRCULATION (i.e. CAROTID PULSE, CAPILLARY FILL, etc.) MUST BE MONITORED.

DO NOT ACTIVATE DEVICE UNTIL READY TO USE. Prolonged exposure to ambient air may yield erroneous results.

Gastric distention, vomitus and/or other fluids or gasses that enter the side arm may cause color change independent of CO₂ concentration.

Indications for Use:


The STAT-Check II is used as an adjunct assessment tool for verification of successful ventilation/intubation. Ventilation, intubation and circulation are confirmed by carefully watching for color change of the CO₂ indicator. The color of the indicator cartridge suggests the percentage of the end-tidal CO₂ in the concentration range of 0 to 5%. The color of the cartridge may be compared to the label to give an approximation of end-tidal CO₂. Lack of color change may indicate improper intubation.

Preparation for Use:

1. Completely tighten the STAT-Check indicator arm onto the threaded connection port by gripping the black protective sleeve and turning clockwise.
2. Remove black protective sleeve from indicator arm.
3. Remove red cap from end of indicator arm if present.

Instructions for Use:

1. Begin ventilation efforts as directed by a medical authority.
2. In the presence of CO₂, the indicator will change from purple to tan or yellow if both cardiac massage and ventilation are being performed correctly.
3. Subsequent color changes from purple to tan or yellow will indicate normal exchange of gases by breathing or resuscitation; compare to label for an approximate assessment of end-tidal CO₂.



CO₂ Indicator Specification

Expiration Date See Indicator Arm
Normal Usage < 4 hours
Optimal Storage Temp. 77°F (25°C)
For Use on patients >2.2 kilograms (5lbs)

Ventlab, LLC
2710 Northridge Dr. NW, Suite A
Grand Rapids, MI 49544 USA

SunMed
Manufactured for Ventlab, LLC,
a Subsidiary of SunMed Holdings
www.Sun-Med.com
Made in China by SunMed AP, LTD



MT Promed Consulting GmbH
Altenhofstrasse 80
66386 St. Ingbert
Germany



SC200.INIFU Rev.1



A.C.E. Stat Check II CO₂ Indicator

Airway Circulation Evaluator



REF SC200

Cautions & Warnings:
THIS DEVICE IS NOT A SUBSTITUTE FOR OBSERVATION OF THE PATIENT. THE STAT-CHECK II MUST NOT BE RELIED UPON AS THE SOLE INDICATOR OF RESUSCITATION EFFECTIVENESS. THIS DEVICE IS AN INDICATOR OF BOTH VENTILATION AND CIRCULATION. THE INDICATOR DYE SHOULD BE CHECKED FREQUENTLY DURING USE. BRAIN DEATH OCCURS IN 4-6 MINUTES WITHOUT OXYGENATED BLOOD FLOW. THE SIGNS OF VENTILATION (i.e. BILATERAL BREATH-SOUND, "FOG SIGN", etc.) and CIRCULATION (i.e. CAROTID PULSE, CAPILLARY FILL, etc.) MUST BE MONITORED.

DO NOT ACTIVATE DEVICE UNTIL READY TO USE. Prolonged exposure to ambient air may yield erroneous results.

Gastric distention, vomitus and/or other fluids or gasses that enter the side arm may cause color change independent of CO₂ concentration.

Indications for Use:


The STAT-Check II is used as an adjunct assessment tool for verification of successful ventilation/intubation. Ventilation, intubation and circulation are confirmed by carefully watching for color change of the CO₂ indicator. The color of the indicator cartridge suggests the percentage of the end-tidal CO₂ in the concentration range of 0 to 5%. The color of the cartridge may be compared to the label to give an approximation of end-tidal CO₂. Lack of color change may indicate improper intubation.

Preparation for Use:

1. Completely tighten the STAT-Check indicator arm onto the threaded connection port by gripping the black protective sleeve and turning clockwise.
2. Remove black protective sleeve from indicator arm.
3. Remove red cap from end of indicator arm if present.

Instructions for Use:

1. Begin ventilation efforts as directed by a medical authority.
2. In the presence of CO₂, the indicator will change from purple to tan or yellow if both cardiac massage and ventilation are being performed correctly.
3. Subsequent color changes from purple to tan or yellow will indicate normal exchange of gases by breathing or resuscitation; compare to label for an approximate assessment of end-tidal CO₂.



CO₂ Indicator Specification

Expiration Date See Indicator Arm
Normal Usage < 4 hours
Optimal Storage Temp. 77°F (25°C)
For Use on patients >2.2 kilograms (5lbs)

Ventlab, LLC
2710 Northridge Dr. NW, Suite A
Grand Rapids, MI 49544 USA

SunMed
Manufactured for Ventlab, LLC,
a Subsidiary of SunMed Holdings
www.Sun-Med.com
Made in China by SunMed AP, LTD



MT Promed Consulting GmbH
Altenhofstrasse 80
66386 St. Ingbert
Germany



SC200.INIFU Rev.1



A.C.E. Stat Check II CO₂ Indicator

Airway Circulation Evaluator



REF SC200

Cautions & Warnings:
THIS DEVICE IS NOT A SUBSTITUTE FOR OBSERVATION OF THE PATIENT. THE STAT-CHECK II MUST NOT BE RELIED UPON AS THE SOLE INDICATOR OF RESUSCITATION EFFECTIVENESS. THIS DEVICE IS AN INDICATOR OF BOTH VENTILATION AND CIRCULATION. THE INDICATOR DYE SHOULD BE CHECKED FREQUENTLY DURING USE. BRAIN DEATH OCCURS IN 4-6 MINUTES WITHOUT OXYGENATED BLOOD FLOW. THE SIGNS OF VENTILATION (i.e. BILATERAL BREATH-SOUND, "FOG SIGN", etc.) and CIRCULATION (i.e. CAROTID PULSE, CAPILLARY FILL, etc.) MUST BE MONITORED.

DO NOT ACTIVATE DEVICE UNTIL READY TO USE. Prolonged exposure to ambient air may yield erroneous results.

Gastric distention, vomitus and/or other fluids or gasses that enter the side arm may cause color change independent of CO₂ concentration.

Indications for Use:


The STAT-Check II is used as an adjunct assessment tool for verification of successful ventilation/intubation. Ventilation, intubation and circulation are confirmed by carefully watching for color change of the CO₂ indicator. The color of the indicator cartridge suggests the percentage of the end-tidal CO₂ in the concentration range of 0 to 5%. The color of the cartridge may be compared to the label to give an approximation of end-tidal CO₂. Lack of color change may indicate improper intubation.

Preparation for Use:

1. Completely tighten the STAT-Check indicator arm onto the threaded connection port by gripping the black protective sleeve and turning clockwise.
2. Remove black protective sleeve from indicator arm.
3. Remove red cap from end of indicator arm if present.

Instructions for Use:

1. Begin ventilation efforts as directed by a medical authority.
2. In the presence of CO₂, the indicator will change from purple to tan or yellow if both cardiac massage and ventilation are being performed correctly.
3. Subsequent color changes from purple to tan or yellow will indicate normal exchange of gases by breathing or resuscitation; compare to label for an approximate assessment of end-tidal CO₂.



CO₂ Indicator Specification

Expiration Date See Indicator Arm
Normal Usage < 4 hours
Optimal Storage Temp. 77°F (25°C)
For Use on patients >2.2 kilograms (5lbs)

Ventlab, LLC
2710 Northridge Dr. NW, Suite A
Grand Rapids, MI 49544 USA

SunMed
Manufactured for Ventlab, LLC,
a Subsidiary of SunMed Holdings
www.Sun-Med.com
Made in China by SunMed AP, LTD



MT Promed Consulting GmbH
Altenhofstrasse 80
66386 St. Ingbert
Germany



SC200.INIFU Rev.1



A.C.E. Stat Check II CO₂ Indicator

Airway Circulation Evaluator



REF SC200

Cautions & Warnings:
THIS DEVICE IS NOT A SUBSTITUTE FOR OBSERVATION OF THE PATIENT. THE STAT-CHECK II MUST NOT BE RELIED UPON AS THE SOLE INDICATOR OF RESUSCITATION EFFECTIVENESS. THIS DEVICE IS AN INDICATOR OF BOTH VENTILATION AND CIRCULATION. THE INDICATOR DYE SHOULD BE CHECKED FREQUENTLY DURING USE. BRAIN DEATH OCCURS IN 4-6 MINUTES WITHOUT OXYGENATED BLOOD FLOW. THE SIGNS OF VENTILATION (i.e. BILATERAL BREATH-SOUND, "FOG SIGN", etc.) and CIRCULATION (i.e. CAROTID PULSE, CAPILLARY FILL, etc.) MUST BE MONITORED.

DO NOT ACTIVATE DEVICE UNTIL READY TO USE. Prolonged exposure to ambient air may yield erroneous results.

Gastric distention, vomitus and/or other fluids or gasses that enter the side arm may cause color change independent of CO₂ concentration.

Indications for Use:

The STAT-Check II is used as an adjunct assessment tool for verification of successful ventilation/intubation. Ventilation, intubation and circulation are confirmed by carefully watching for color change of the CO₂ indicator. The color of the indicator cartridge suggests the percentage of the end-tidal CO₂ in the concentration range of 0 to 5%. The color of the cartridge may be compared to the label to give an approximation of end-tidal CO₂. Lack of color change may indicate improper intubation.

Preparation for Use:

1. Completely tighten the STAT-Check indicator arm onto the threaded connection port by gripping the black protective sleeve and turning clockwise.
2. Remove black protective sleeve from indicator arm.
3. Remove red cap from end of indicator arm if present.

Instructions for Use:

1. Begin ventilation efforts as directed by a medical authority.
2. In the presence of CO₂, the indicator will change from purple to tan or yellow if both cardiac massage and ventilation are being performed correctly.
3. Subsequent color changes from purple to tan or yellow will indicate normal exchange of gases by breathing or resuscitation; compare to label for an approximate assessment of end-tidal CO₂.



CO₂ Indicator Specification

Expiration Date See Indicator Arm
Normal Usage < 4 hours
Optimal Storage Temp. 77°F (25°C)
For Use on patients >2.2 kilograms (5lbs)

Ventlab, LLC
2710 Northridge Dr. NW, Suite A
Grand Rapids, MI 49544 USA

SunMed
Manufactured for Ventlab, LLC,
a Subsidiary of SunMed Holdings
www.Sun-Med.com
Made in China by SunMed AP, LTD



MT Promed Consulting GmbH
Altenhofstrasse 80
66386 St. Ingbert
Germany



SC200.INIFU Rev.1