



Top Valve Vent Mask™

Premium Inflatable Non-Conductive Anesthesia Mask

REF VR4500 / 4 / Small Adult

Directions for Use

Inspect cushion for proper inflation pressure before use. Use luer tip syringe to adjust cushion pressure, if needed.



Caution

Do not attempt to sterilize, autoclave or clean this device.



Single Use



Ventlab, LLC
2710 Northridge Dr. NW, Suite A
Grand Rapids, MI 49544 USA



Manufactured for Ventlab, LLC,
a Subsidiary of SunMed Holdings, LLC
www.Sun-Med.com
Made in China by SunMed AP, LTD

VR4500.IFU Rev.3 2022/11